



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

HEALTH / ALCOHOL PROGRAM

11-1-87

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN
204108

DATE OF INSPECTION

5-28-09

LOCATION OF INSTRUMENT (STREET AND CITY)

Ralls Co. Sheriffs Dept., 301 Ralls St., New London, MO 63459

TIME OF INSPECTION

2336

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER 49 °C

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C)

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 0.097

TEST 2 0.097

TEST 3 0.098

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS: (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS 0 (0-.04) 0 (.05-.09) 1 (.10-.14) 0 (.15-.19) 0 (Over .19) 0

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

Instrument meets Dept. of Health specs., Guth Labs., 10 solution, lot no. 08340, exp. 10-15-09

INSPECTING OFFICER

SIGNATURE

B. C. Griffin

PRINT NAME

B. C. Griffin

TYPE II PERMIT NUMBER/EXPIRATION DATE

920080 / 04-22-2011

TELEPHONE NUMBER

(660) 385-2132



GUTH LABORATORIES, INC.

590 NORTH 67th STREET • HARRISBURG, PA 17111-4511 • TELEPHONE: 717-564-5470

CERTIFICATE OF ANALYSIS

Certified Alcohol Reference Solution for Simulator

Random Samples of Lot Number **08340** of Alcohol Reference Solution for Simulator were analyzed by gas chromatography and found to contain **0.1211** percent (w/vol) ethyl alcohol. The expiration date for this lot number is **October 15, 2009** at **11:59 PM**.

When used in a calibrated Simulator, operating at **34°C +/- .2°C**, this solution will give a breath alcohol analysis instrument reading of **0.10** percent BAC.

The alcohol and water used in this solution were free of test interfering substances.

Ted L. Pauley, President
GUTH LABORATORIES, INC.

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204108
05/28/09
23:36


--- DIAGNOSTIC CHECK ---

COMPUTER:	OKAY
PROGRAM:	OKAY
HEATERS	
SAMPLE CHAMBER:	49c
FLOW DETECTOR:	OKAY
PUMP	
HIGH SPEED:	OKAY
DETECTOR:	OKAY
FILTERS:	OKAY
QUARTZ STANDARD:	OKAY
CALIBRATION:	OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefgijklmnop
qrstuvwxyz{|}~

Operator Signature



Printed on recycled paper with agri-based inks

CMSU 2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204108
05/28/09

ARREST TIME: 23:00
SUBJECT NAME:
TEST
DOB: 01/01/01 SEX: M
STATE/D.L.: MO/1234567890
ARRESTING OFFICER:
GRIFFIN/B/C
OFFICER I.D.: 0429
TESTING OFFICER:
GRIFFIN/B/C
OFFICER I.D.: 0429
PERMIT NUMBER: 920080
EXPIRATION DATE: 04/22/11
MISCELLANEOUS DATA:

--- BREATH ANALYSIS ---

BLANK TEST	.000	23:57
INTERNAL STANDARD	VERIFIED	23:57
RADIO INTERFERENCE		

Operator Signature

Printed on recycled paper with agri-based inks

CMSU

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC DATAMASTER SERIAL NUMBER 204108
05/28/09

TESTING OFFICER:
GRIFFIN/B/C
OFFICER I.D.: 0429
PERMIT NUMBER: 920080
EXPIRATION DATE: 04/22/11
MISCELLANEOUS DATA:

--- SUPERVISOR MODE ---

BLANK TEST	.000	23:46
INTERNAL STANDARD	VERIFIED	23:46
EXTERNAL STANDARD	.097	23:47
BLANK TEST	.000	23:48
EXTERNAL STANDARD	.097	23:48
BLANK TEST	.000	23:49
EXTERNAL STANDARD	.098	23:50
BLANK TEST	.000	23:50

N = 3
SIM. = .1
AVG. = .0973

Operator Signature

Printed on recycled paper with agri-based inks

CMSU 2208-02

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



BRANDON GRIFFIN

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 04/22/09

Number 920080

Expires 04/22/2011

MO 580-0771 (7-88)

John J. Mathewson

Director of State Public Health Laboratory

Margaret T. Donnelly

Director, Department of Health

Lab. 4 (R7-88)